

## 5P1Transducer



**LEGAL MANUFACTURER**

**BK Medical**  
Mileparken 34  
2730 Herlev  
Denmark  
Tel.:+45 4452 8100/Fax:+45 4452 8199  
[www.bkultrasound.com](http://www.bkultrasound.com)  
Email: [info@bkultrasound.com](mailto:info@bkultrasound.com)

The serial number on a BK Medical product contains information about the year of manufacture.

**BK Medical Customer Satisfaction**

Input from our customers helps us improve our products and services. Your opinions are important to us. You are always welcome to contact us via your BK Medical representative or by contacting us directly.

**5P1 = Ref. Type 9077**

© 2018 BK Medical

Information in this document may be subject to change without notice.

# Contents


---

- Introduction ..... 5
  - Intended use ..... 5
  - Indications for Use ..... 5
- General Information ..... 6
  - Service and Repair ..... 6
  - Caring for the Transducer ..... 6
- Reprocessing ..... 6
- Starting Imaging ..... 7
  - Connecting the Transducer ..... 7
  - Changing Frequency ..... 7
  - Using a Transducer Cover ..... 7
  - Using the Transducer Control Button ..... 8
  - Changing Orientation ..... 8
- Cleaning after use ..... 9
- Disposal ..... 9



## Introduction

This is the user guide for the 5P1 transducer, and it must be used together with the *Care and Cleaning* user guide, which contains important safety information.

 Physicians only	<b>Caution</b> Rx-c1 United States Federal law restricts this device to sale by or on the order of a physician.
---	--

### Intended use

The transducer is intended for diagnostic ultrasound imaging or fluid flow analysis of the human body

### Indications for Use

5P1 is designed for basic cardiac adult examinations (including Focus Assessed Transthoracic Echo [FATE] and Focused Assessment with Sonography for Trauma [FAST]), abdominal imaging and adult cephalic (transcranial) imaging and CW Doppler imaging.

### Patient Population

The patient population is adults.




Figure 1. 5P1 transducer


### Contraindication

Cardiac Adult is not for direct use on the heart.


## General Information

Product specifications, acoustic output data and data about EMC (electromagnetic compatibility) for this transducer can be found in the *Product Data Sheet* and the *Technical Data (BZ2100)* that accompany this user guide.

	<p><b>WARNING</b> GS-w2</p> <p>If at any time the system malfunctions, or the image is severely distorted or degraded, or you suspect in any way that the system is not functioning correctly:</p> <ul style="list-style-type: none"><li>• Remove all transducers from contact with the patient.</li><li>• Turn off the system. Unplug the system from the wall and make sure it cannot be used until it has been checked.</li><li>• Do not try to repair the system yourself.</li><li>• Contact your BK service representative or hospital technician.</li></ul>
---	---

	<p><b>WARNING</b> AO-w1</p> <p>To avoid tissue damage, always keep the exposure level (the acoustic output level and the exposure time) as low as possible.</p>
---	---

## Service and Repair

	<p><b>WARNING</b> SR-w1</p> <p>Service and repair of BK electromedical equipment must be carried out only by the manufacturer or its authorized representatives. BK Medical reserves the right to disclaim all responsibility, including but not limited to responsibility for the operating safety, reliability and performance of equipment serviced or repaired by other parties. After service or repairs have been carried out, a qualified electrician or hospital technician should verify the safety of all equipment.</p>
---	--

## Caring for the Transducer


The transducer may be damaged during use or reprocessing, so it must be checked before use for cracks or irregularities in the surface, following the procedure in *Care and Cleaning*. It should also be checked thoroughly once a month following the same procedure.

## Reprocessing

To ensure the best results when using BK Medical equipment, it is important to maintain a strict cleaning routine.


Complete details and procedures can be found in *Care and Cleaning*, which accompanies this user guide. A list of reprocessing methods that the transducer can withstand are listed in the *Product Data Sheet*.

Sterile cover is available. See the *Product Data Sheet* for more information.


	<p><b>WARNING</b> Reproc-w2</p> <p>Users of this equipment have an obligation and responsibility to provide the highest possible degree of infection control to patients, co-workers and themselves. The instructions in this book are meant as a guide. To avoid cross-contamination, follow all infection control policies (including for reprocessing, packing and storage) for personnel and equipment that have been established for your office, department or hospital.</p>
---	--

## Starting Imaging

Before use, all equipment must be reprocessed according to expected use.

	<p><b>WARNING</b> T-w5</p> <p>To prevent electrical shock and damage to the transducer, the connector pins in the transducer plug must always be completely dry before you connect to a system.</p>
---	---

## Connecting the Transducer

	<p><b>WARNING</b> GS-w4a</p> <p>It is essential for the patient's safety that only the correct equipment is used.</p> <ul style="list-style-type: none"> <li>• Do not use other manufacturers' transducers with BK systems.</li> <li>• Do not use BK transducers with other manufacturers' systems.</li> <li>• Do not use unauthorized combinations of transducers and needle guides.</li> </ul>
---	--

The transducer is connected to the system using the array transducer socket on the system. To connect, flip the system's locking lever to the right. Align the transducer plug to the system socket and insert securely. Flip the system's locking lever to the left to lock it.

When connected, the transducer complies with Type BF requirements of EN60601-1 (IEC 60601-1).

## Changing Frequency

The multifrequency Imaging (MFI) facility enables you to select the imaging frequency. See the applicable system user guide for instructions.

## Using a Transducer Cover

BK recommends the use of a sterile transducer cover to reduce the risk of cross-contamination. See the *Product Data Sheet* for a list of available transducer covers. Follow local guidelines for the use of transducer covers in your area.

**NOTE:** *In the United States of America, it is recommended to use transducer covers that have been market cleared. In Canada, use only licensed transducer covers. In Europe, transducer covers must be CE-marked.*

**WARNING TC-w1**

Some transducer covers can contain latex. Because of reports of severe allergic reactions to medical devices containing latex (natural rubber), the FDA advises health-care professionals to identify their latex-sensitive patients and be prepared to treat allergic reactions promptly.

Apply sterile gel to the tip of the transducer or fill the cover with 1 to 2 ml of sterile water. This improves screen imaging by preventing image artifacts caused by air bubbles.

Gel also creates a good acoustic contact between the skin and the transducer; therefore, apply a small amount to the outside of the cover prior to imaging and re-apply frequently.

Follow these precautions when putting sterile covers on a transducer:

- Wear sterile gloves.
- When using a puncture attachment, place it gently over the cover and secure it, following the instructions for the puncture attachment.
- Verify that the cover has not been damaged in the process. If it has, repeat the procedure with a new transducer cover.

**Caution T-c3**

Use only water-based gel (sterile if you are using a sterile transducer cover). Products containing parabens, petroleum, or mineral oils may harm the transducer or transducer cover.

## Using the Transducer Control Button

The control button on the transducer controls the imaging.

Press the button to **Start** or **Stop** imaging (freeze frame). Press the button for more than one second to make a copy of the image.


The transducer makes a “beep” sound each time you press the button.

## Changing Orientation

To change the orientation of the image on the monitor, refer to the applicable system user guide for instructions.




## Cleaning after use

	<p><b>WARNING</b> Reproc-w3</p> <p>Immediately after use, you must pre-clean the device until visually clean (including device lumens if existing). Conduct the thorough cleaning process as soon as possible after use in order to prevent bioburden drying on the surface. Dried bioburden can lead to inefficient cleaning, disinfection and sterilization, causing a risk of cross-contamination.</p> <p>If pre- and thorough cleaning cannot be done immediately, keep the device moist until cleaning.</p>
---	--

Use a suitable brush to make sure that biological material and gel are removed from all channels and grooves. See *Care and Cleaning* for cleaning instructions.

## Disposal

When the transducer is scrapped at the end of its life, national rules for the relevant material in each individual land must be followed. Within the EU, when you discard the transducer, you must send it to appropriate facilities for recovery and recycling.

	<p><b>WARNING</b> D-w1</p> <p>For disposal of contaminated items such as transducer covers or needle guides or other disposable items, follow disposal control policies established for your office, department or hospital.</p>
--	--





 BK Medical ApS, Mileparken 34, 2730 Herlev, Denmark. T +45 4452 8100 F +45 4452 8199



powered by analogic 

**BK Ultrasound**  
8 Centennial Drive  
Peabody  
MA 01960  
USA  
T +1 978-326-1300  
bkultrasound.com

**USA  
Sales & Service**  
BK Ultrasound  
8 Centennial Drive  
Peabody  
MA 01960  
USA  
T +1 978-326-1300  
F +1 978-326-1399  
bkultrasound.com

**Europe and Rest of World  
Sales, Service & Design Center**  
BK Ultrasound  
Mileparken 34  
2730 Herlev  
Denmark  
T +45 4452 8100  
F +45 4452 8199  
bkultrasound.com

**Asia  
Sales & Service**  
Analogic Medical Equipment  
(Shanghai) Co., Ltd.  
1377, Lan Dian Road  
Pu Dong New District  
Shanghai  
China 201132  
T +86 21 2089 0333  
bkultrasound.com